



An ISO 9001 & 14001 Co.

# MEDICAL & HOSPITAL EQUIPMENT



## MODEL NO.:- RAJAS-CT5-19 (Resuscitation kit)

### Specification:-

#### Features :

- Insufflator

Transparent self-filling with one-way valve

Approximately 312 ml

Insufflation approximately 145 ml

Intentional leak device

- Ambient air inlet device
- Patient device

System for stopping the phenomenon of inhalation of exhaled air, propagating in one direction only, newborn model

Higher than normal pressure relief device

Transparent engineering plastic PSU

Comes with clear silicone masks #0 and #1, fitting 2.5 / 1.5 cm eartips

- Adjustable device between 1.5 and 20 cm H2O
- Easy disassembly for removal of unwanted substances
- All elements are sterilizable at approx. 132 °C
- Not free from microorganisms

Delivery will be made to the following sites DSF 4 (OUAGADOUGOU and 32 by DRS center -south (Manga) DRS plateau-central (Ziniaré), DRS Center-north (KAYA), DRS EST (Fada) DRS CASCADES (Banfora) DRS north Ouahigouya), DRS CENTRE-EST (Tenkodogo), DRS Boucle du Mouhoun (Dédougou) DRS Sahel (DORI) South-west (Gaoua)

Delivery will be made to the following sites 1 by DRS DRS center-south (Manga) DRS plateau-central (Ziniaré), DRS Center-north (KAYA), DRS EST (Fada) DRS CASCADES (Banfora) DRS north Ouahigouya), DRS CENTRE-EST (Tenkodogo), DRS Boucle du Mouhoun (Dédougou) DRS Sahel (DORI) South-west (Gaoua)

# RAJAS ENTERPRISES (INDIA)

#59, Phase-1, Sector-1, HSIIDC, Industrial Estate, SAHA, Ambala Cantt-133104 (India)

Email:rajasmmedical@gmail.com, rajasglass@gmail.com

www.rajasesenterprises.com