



An ISO 9001 & 14001 Co.

MEDICAL & HOSPITAL EQUIPMENT



MODEL NO.:- RAJAS-CT5-04
(Pediatric electric aspirator)

Specification:-

Pediatric electric aspirator

- Continuous suction and pedal type

Waterproof pneumatic

- With vacuum adjustment knob
- Depression reaching 680 mbar or more
- On/off switch
- 2 autoclavable 2-litre jars, one of which is

Service and the other in reserve

- Flow rate 60L/min or more
- Overflow device (filling safety)
- Power supply 220 to 240V 50Hz
- To be delivered and installed complete with all Accessories (jars, 2 suction hoses 1.5 m long) Each, 2 suction cannulas), manual Use and maintenance instructions in French.

Delivery will be made to the following sites DSF 2, Ouagadougou and 15 by DRS DRS center-south (Manga) DRS plateau-central (Ziniaré), DRS Center-north (KAYA), DRS EST (Fada) DRS CASCADES (Banfora) DRS north Ouahigouya), DRS CENTRE-EST (Tenkodogo), DRS Boucle du Mouhoun (Dédougou) DRS Center Ouagadougou, DRS Hauts-Bassins (Bobo) DRS Sahel (DORI) South-west (Gaoua)

RAJAS ENTERPRISES (INDIA)

#59, Phase-1, Sector-1, HSIIDC, Industrial Estate, SAHA, Ambala Cantt-133104 (India)

Email:rajasmmedical@gmail.com, rajasglass@gmail.com

www.rajasenterprises.com