



An ISO 9001 & 14001 Co.

MEDICAL & HOSPITAL EQUIPMENT



MODEL NO.:- RAJAS-CT3-08
(liter pressure cooker)

Specification:-

Steam sterilizer of the pressure cooker type intended for the destruction by hydrolysis of bacteria for the sterilization of contaminated equipment and instruments. Tank and casing in stainless steel; capacity of at least 11 liters; heating by gas stove or electricity; built-in control pressure gauge; pressure fixing device; automatic discharge with cold air

RAJAS ENTERPRISES (INDIA)

#59, Phase-1, Sector-1, HSIIDC, Industrial Estate, SAHA, Ambala Cantt-133104 (India)
Email: rajasmedical@gmail.com, rajasglass@gmail.com
www.rajasenterprises.com